# Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

| MANUFACTURER | | |
|---|---|---|
| **Name of Company** | **Address** | **SRN** |
| AdDent, Inc. | 43 Miry Brook Road<br>Danbury, CT 06810 USA | US-MF-000019453 |

| AUTHORIZED REPRESENTATIVE | | | |
|---|---|---|---|
| **Name of Company** | **Address** | **SRN** | **Phone/email** |
| Emergo Europe | Prinsessegracht 20<br>2514 AP The Hague<br>The Netherlands | NL-AR-000000116 | +31.70.345.8570<br>EmergoEurope@ul.com |

| PRODUCT IDENTIFICATION | |
|---|---|
| **Product Name** | **Code / Catalog Number** |
| Microlux 3 mm Light Guide<br>Microlux 2mm Light Guide<br>Microlux 2mm Pedo Light Guide | 630003<br>600045<br>110066 |
| **Intended Purpose** | **Basic UDI-DI** |
| The Microlux 3mm, 2mm and 2mm Pedo Light Guides are accessories for the Microlux family of Transilluminator units. They are intended to be connected to a Microlux Transilluminator and used by a dentist or dental professional as an adjunct to traditional oral examination. They are used for trans-illumination of the teeth, using high intensity LED light, thus aiding clinical dental examinations. Trans-illumination helps to visualize crown fractures, caries and root canal orifice. | ++D869633000B6 |

| RISK CLASS FOR DEVICES | | |
|---|---|---|
| **Device Classification** | | **Common Specifications / Standards** |
| **Class:** | I | ISO 9001:2015 certified, ISO 13485:2016 compliant, EN ISO 14971:2012 |
| **Rule:** | 5 | |

AdDent, Inc. declares that the above-mentioned products meet the provision of the following EU legislation:

- Medical Devices Regulation (EU) 2017/745
- RoHs (EU) 2015/863

**COMPANY REPRESENTATIVE:** Dr. Joshua Friedman

**TITLE:** President

**SIGNATURE:**



**PLACE:** Danbury, CT USA

**DATE:** August 1, 2022